

Certificate of Analysis

Product Name: Pesticides Calibration Standard

Product Number: US-325-1

Lot Issue Date: 23-Feb-2021

Lot Number: 0006589724

Expiration Date: 31-Mar-2023

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration ± Uncertainty
chlorobenzilate	000510-15-6	RM17349	1005 ± 5 µg/mL
chloroneb	002675-77-6	RM00198	1001 ± 5 µg/mL
chlorothalonil	001897-45-6	RM15959	1001 ± 5 µg/mL
DCPA	001861-32-1	NT01783	1001 ± 5 µg/mL
etridiazole	002593-15-9	RM16064	1001 ± 5 µg/mL
hexachlorobenzene	000118-74-1	RM03125	1002 ± 5 µg/mL
cyanazine	021725-46-2	RM15773	1003 ± 5 µg/mL
cis-permethrin	061949-76-6	RM16393	1000 ± 5 µg/mL
trans-permethrin	061949-77-7	RM16541	1001 ± 5 µg/mL
propachlor	001918-16-7	RM16979	1003 ± 5 µg/mL
trifluralin	001582-09-8	RM09256	1002 ± 5 µg/mL
trans-nonachlor	039765-80-5	RM03191	1000 ± 5 µg/mL
cis-chlordane	005103-71-9	RM07243	1001 ± 5 µg/mL
trans-chlordane	005103-74-2	RM02726	1004 ± 5 µg/mL
chlorpyrifos	002921-88-2	RM17484	1001 ± 5 µg/mL

Matrix: methyl tert-butyl ether (MTBE)

Storage Conditions: Store at Room Temperature (15° to 30°C).



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality Management System. Cert # 56 100 18560026

Page: 1 of 2

www.agilent.com/quality/
CSD-QA-015.1



ISO 17025 Cert
No. AT-1937

Certificate of Analysis

Product Number: US-325-1

Lot Number: 0006589724

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification. □

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.

Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:



Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality Management System. Cert # 56 100 18560026

Page: 2 of 2

www.agilent.com/quality/
CSD-QA-015.1



ISO 17025 Cert
No. AT-1937